

Reference No. : WTH22H09191213E
 Applicant (Holder) : ShenZhen RiHuiDa Power Supply Co., Ltd.
 2,3 & 4F,NO. B Building, Rihuida Technology park,No.28 Hongyuan Road,
 Address : Pingdong Community,Pingdi Sub-District.Longgang District.Shenzhen.
 PRC.
 Manufacturer : ShenZhen RiHuiDa Power Supply Co., Ltd.
 2,3 & 4F,NO. B Building, Rihuida Technology park,No.28 Hongyuan Road,
 Address : Pingdong Community,Pingdi Sub-District.Longgang District.Shenzhen.
 PRC.
 Product Name : Power adapter
 Trade Mark : N/A
 Model(s) : RHDzzWxxxyyy
 Rating : Input: AC100-240V, 50/60Hz, 48W
 Output: DC3-48V, 0.1-6A, 48W

Test Standards:

EN 60601-1-2:2015+A1:2021, EN IEC 61000-3-2:2019+A1:2021, EN 61000-3-3:2013+A2:2021.

The above product has been tested by us with the listed standards and found in compliance with the 2014/30/EU European Electromagnetic Compatibility.

It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

EN 60601-1-2: Medical electrical equipment - Part 1-2:General requirements for basic safety and essential performance - Collateral Standard:Electromagnetic disturbances - Requirements and tests

EN IEC 61000-3-2: Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions (equipment input current ≤ 16 A per phase)

EN 61000-3-3: Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the above mentioned EU Directive. Other relevant Directives have to be observed.

After preparation of the necessary technical documentation as well as the conformity declaration, the CE marking as shown below can be affixed on the equipment.



Galt, Ryan / Project Manager
 Date: January 13, 2023

The statement is based on a single evaluation of the sample of above mentioned product(s). It does not imply an assessment of the whole production.

Shenzhen Hongcai Testing Technology Co., Ltd.

Tel: 0755-84616666

Http://www.hongcai@hct-test.com

Service Tel: 400-0066-989

E-mail: service@hct-test.com