



CERTIFICATE OF PRODUCT CERTIFICATION

CERTIFICATE NO.: CQC25001476866

Valid from: Jul.01,2025

Valid until:Jun.30,2030

NAME AND REGISTERED ADDRESS OF THE APPLICANT ShenZhen RiHuiDa Power Supply Co., Ltd.
301,401,501, No.28 Hongyuan Road, Pingdong Community, Pingdi Sub-District. Longgang District.
ShenZhen, PRC.

NAME AND REGISTERED ADDRESS OF THE MANUFACTURER ShenZhen RiHuiDa Power Supply Co., Ltd.
301,401,501, No.28 Hongyuan Road, Pingdong Community, Pingdi Sub-District. Longgang District.
ShenZhen, PRC.

NAME AND LOCATION OF THE FACTORY ShenZhen RiHuiDa Power Supply Co., Ltd.
301,401,501, No.28 Hongyuan Road, Pingdong Community, Pingdi Sub-District. Longgang District.
ShenZhen, PRC.

PRODUCT NAME, MODEL AND SPECIFICATION Medical-grade power adapter
See Appendix (Altitude up to 5000m) (Sale without cord set)

THE STANDARDS AND TECHNICAL REQUIREMENTS FOR THE PRODUCTS GB 9706.1-2020;YY 9706.102-2021

TYPE OF CERTIFICATION SCHEMES Type testing+Follow up factory inspection

This is to certify that the above mentioned product(s) complies with the requirements of certification rules of CQC11-461271-2023.
The validity of the certificate is subject to positive result of the regular follow up inspection by issuing certification body until the expiry date.

The certificate information is available through the QR code below or CNCA's website: www.cnca.gov.cn



SIGNATURE:

谢肇煦



中国质量认证中心
CHINA QUALITY CERTIFICATION CENTRE

CHINA QUALITY CERTIFICATION CENTRE CO., LTD.
Section 9, No. 188, Nansihuan Xilu, Beijing 100070 P. R. China

Tel: +86 10 83886666
<http://www.cqc.com.cn>